

TruSight™ Tumor 170

A comprehensive next-generation sequencing assay that targets DNA and RNA variants from the same formalin-fixed, paraffin-embedded (FFPE) tumor sample.

Highlights

- Comprehensive Coverage of Cancer-Related Variants**
 Single-assay efficiency using DNA and RNA for assessment of small variants, amplifications, splice variants, and fusions
- Integrated, Streamlined Workflow**
 DNA and RNA libraries are prepared in parallel with an integrated workflow following DNA shearing/cDNA synthesis
- Accurate Results from Low-Quality Samples**
 Variant detection with 40 ng DNA/RNA input, and as low as 5% mutant allele frequency, from FFPE samples

To help researchers address this challenge, Illumina offers TruSight Tumor 170, a next-generation sequencing (NGS) assay designed to cover 170 genes associated with solid tumors. TruSight Tumor 170 is an enrichment-based targeted panel that simultaneously analyzes DNA and RNA, covering a wide range of genes and variant types. The panel is designed to work with the NextSeq™ 500, NextSeq 550, or HiSeq™ 2500 Sequencing Systems (Figure 1).

Comprehensive Cancer-Related Content Design

TruSight Tumor 170 targets all coding exons, per the current RefSeq database,² in 170 genes (Table 1). The genes and type of variant analysis for each gene were carefully selected to include content cited by professional organizations such as the National Comprehensive Cancer Network (NCCN) and the European Society for Medical Oncology (ESMO).^{3,4} Independent consortia publications and late-stage pharmaceutical research also influenced the design of TruSight Tumor 170. The content includes 55 genes for fusions and splice variants, 148 SNVs and indels, and 59 amplifications. By harnessing the expertise of recognized authorities in the oncology community, TruSight Tumor 170 provides researchers with comprehensive coverage of the variants that are most likely to play a role in tumorigenesis.

Introduction

Cancer is a leading cause of death worldwide and has the potential to originate in any tissue.¹ Analyzing the genetic basis of a given tumor is important for understanding its progression and developing new methods of treatment. However, numerous genes can cause or influence tumor progression, and many heterogeneous tumors carry multiple mutations. Furthermore, the function of any gene can be altered by several types of variations including single nucleotide variants (SNVs), multiple nucleotide variants (MNVs), small insertions or deletions (indels), amplifications, splice variations, and gene fusions. Therefore, it is difficult for researchers to analyze tumors efficiently when available methods only cover a portion of these variations, and sequential testing consumes valuable tissue, time, and resources.




Figure 1: TruSight Tumor 170 Workflow—TruSight Tumor 170 is optimized for integration into current lab workflows, going from extracted nucleic acids to variant calling in less than 4 days. The assay can be run on the NextSeq Series or HiSeq 2500 System.

Table 1: Gene Content in the TruSight Tumor 170 Assay

SNVs and Indels (from DNA)									
AKT1	BRIP1	CREBBP	FANCI	FGFR2	JAK3	MSH3	PALB2	RAD51D	TSC1
AKT2	BTK	CSF1R	FANCL	FGFR3	KDR	MSH6	PDGFRA	RAD54L	TSC2
AKT3	CARD11	CTNNB1	FBXW7	FGFR4	KIT	MTOR	PDGFRB	RB1	VHL
ALK	CCND1	DDR2	FGF1	FLT1	KMT2A (MLL)	MUTYH	PIK3CA	RET	XRCC2
APC	CCND2	DNMT3A	FGF2	FLT3	KRAS	MYC	PIK3CB	RICTOR	
AR	CCNE1	EGFR	FGF3	FOXL2	MAP2K1	MYCL1	PIK3CD	ROS1	
ARID1A	CD79A	EP300	FGF4	GEN1	MAP2K2	MYCN	PIK3CG	RPS6KB1	
ATM	CD79B	ERBB2	FGF5	GNA11	MCL1	MYD88	PIK3R1	SLX4	
ATR	CDH1	ERBB3	FGF6	GNAQ	MDM2	NBN	PMS2	SMAD4	
BAP1	CDK12	ERBB4	FGF7	GNAS	MDM4	NF1	PPP2R2A	SMARCB1	
BARD1	CDK4	ERCC1	FGF8	HNF1A	MET	NOTCH1	PTCH1	SMO	
BCL2	CDK6	ERCC2	FGF9	HRAS	MLH1	NOTCH2	PTEN	SRC	
BCL6	CDKN2A	ERG	FGF10	IDH1	MLLT3	NOTCH3	PTPN11	STK11	
BRAF	CEBPA	ESR1	FGF14	IDH2	MPL	NPM1	RAD51	TERT	
BRCA1	CHEK1	EZH2	FGF23	INPP4B	MRE11A	NRAS	RAD51B	TET2	
BRCA2	CHEK2	FAM175A	FGFR1	JAK2	MSH2	NRG1	RAD51C	TP53	
Amplifications (from DNA)									
AKT2	BRCA2	CHEK1	ERCC2	FGF5	FGF14	FGFR4	MDM4	NRG1	RAF1
ALK	CCND1	CHEK2	ESR1	FGF6	FGF19	JAK2	MET	PDGFRA	RET
AR	CCND3	EGFR	FGF1	FGF7	FGF23	KIT	MYC	PDGFRB	RICTOR
ATM	CCNE1	ERBB2	FGF2	FGF8	FGFR1	KRAS	MYCL1	PIK3CA	RPS6KB1
BRAF	CDK4	ERBB3	FGF3	FGF9	FGFR2	LAMP1	MYCN	PIK3CB	TFRC
BRCA1	CDK6	ERCC1	FGF4	FGF10	FGFR3	MDM2	NRAS	PTEN	
Fusions and Splice Variants (from RNA)									
ABL1	BRAF	EML4	ETV4	FGFR4	KIF5B	MYC	NTRK2	PIK3CA	TMPRSS2
AKT3	BRCA1	ERBB2	ETV5	FLI1	KIT	NOTCH1	NTRK3	PPARG	
ALK	BRCA2	ERG	EWSR1	FLT1	KMT2A (MLL)	NOTCH2	PAX3	RAF1	
AR	CDK4	ESR1	FGFR1	FLT3	MET	NOTCH3	PAX7	RET	
AXL	CSF1R	ETS1	FGFR2	JAK2	MLLT3	NRG1	PDGFRA	ROS1	
BCL2	EGFR	ETV1	FGFR3	KDR	MSH2	NTRK1	PDGFRB	RPS6KB1	

Combined Workflow for DNA and RNA

Library preparation for TruSight Tumor 170 uses an enrichment method that can be simultaneously applied to DNA and RNA extracted from the same sample. After the initial steps, in which genomic DNA is sheared and RNA is converted to cDNA, library prep becomes a combined workflow (Figure 2).

- Sheared DNA and cDNA are converted into sequenceable libraries.
- Regions of interest are hybridized to biotinylated probes, magnetically pulled down with streptavidin-coated beads, and eluted to enrich the library pool.
- Libraries are normalized using a simple bead-based protocol before pooling and sequencing.

TruSight Tumor 170 Data Analysis

Illumina sequencing systems offer the option to connect to BaseSpace® Sequence Hub, the Illumina genomics computing environment for sequencing data analysis and management. Researchers can securely store, analyze, archive, and share sequencing data. The TruSight Tumor 170 App is designed to make variant calls that enable downstream reporting in an easy-to-read

format. Raw data outputs for small variants, amplifications, fusions, and splice variants are provided, as well as user-friendly, focused outputs for high confidence RNA variants and fusion results.

The TruSight Tumor 170 App is available in BaseSpace Sequence Hub. For users who desire locally based secondary analysis, Illumina offers a Docker-based image of the app. Contact your sales or support representative for further information.

Sensitive, Highly Confident Variant Detection

Deep sequencing using NGS provides the high sensitivity to reveal somatic variation in tumor subpopulations. Illumina sequencing by synthesis (SBS) chemistry is the most widely adopted NGS technology, generating > 90% of global sequencing data.* When paired with high-quality sequencing on the NextSeq and HiSeq Systems, TruSight Tumor 170 provides uniform coverage of target regions, identifying somatic mutations as low as 5% mutant allele frequency with ≥ 250x minimum coverage (Table 2).



Figure 2: Combined Library Prep Workflow—The DNA and RNA samples follow the same workflow, after the cDNA synthesis step (for RNA) and the shearing step (for DNA).

Table 2: Specifications

Parameter	Details
System	NextSeq or HiSeq 2500 System
Panel Size	533 kb DNA 358 kb RNA
Minimum Insert Size	79 bp DNA 63 bp RNA
DNA Input Requirement	40 ng total
RNA Input Requirement	40 ng total
Library Preparation Time	32 hours
Sequence Run Time	24 hours (NextSeq Systems) or 27 hours (HiSeq 2500 System)
Sequence Run	2 × 101 cycles
Kit Size	24 samples (both DNA and RNA)
Sample Throughput	8 samples per run (NextSeq Systems) or 6 samples per rapid run (HiSeq 2500 System)
Sensitivity	5% Mutant Allele Frequency > 95% sensitivity and specificity

High Coverage of Targets from Low-Quality Samples

Nucleic acids extracted from FFPE tissues have the potential to fail quality control checks and yield poor target coverage resulting in low analytical sensitivity. TruSight Tumor 170 addresses this issue by generating libraries from nucleic acids of small fragment size, as low as 79 bp for DNA and 63 bp for RNA. This enables deep coverage of FFPE samples, even when the quality of extracted nucleic acids is low (Figure 3).




Figure 3: Target Coverage from FFPE Samples—DNA from FFPE tumor samples of varying quality was extracted and evaluated using the TruSight Tumor 170 assay and sequenced on the NextSeq 500 System. Quality of each sample was also assessed using qPCR to measure DNA amplification potential. The ΔCq value indicates the cycle threshold (Ct) value of each DNA sample minus the Ct value of a DNA standard.

* Data Calculations on file, Illumina, Inc., 2015.

Reliable Small Variant Detection From High and Low-Quality Samples

TruSight Tumor 170 provides sensitivity and accuracy for identifying low-frequency variations in samples of varying quality. High target coverage enables confident calling of low-level variants in characterized cell lines (Table 3). TruSight Tumor 170 enables variant detection in FFPE tumor samples with as low as 5% mutant allele frequency (Table 4).

Table 3: Small Variant Calling with Characterized Cell Lines

Gene	Mutation	Reported Frequency	Detected Frequency	Coverage
APC	R2714C	0.33	0.31	2547x
ARID1A	P1562fs	0.34	0.31	419x
BRAF	V600E	0.10	0.11	2282x
BRCA2	A1689fs	0.33	0.30	1097x
EGFR	G719S	0.24	0.22	2207x
EP300	K291fs	0.08	0.06	1359x
FBXW7	G667fs	0.34	0.30	2870x
FGFR1	P150L	0.08	0.08	1102x
FLT3	S985fs	0.10	0.10	1925x
FLT3	V197A	0.12	0.10	1908x
IDH1	S261L	0.10	0.09	2052x
KIT	D816V	0.10	0.15	1239x
KRAS	G13D	0.15	0.14	1507x
KRAS	G12D	0.06	0.07	1503x
MET	V237fs	0.06	0.06	3700x
MLH1	L323M	0.08	0.09	1725x
NF1	L626fs	0.08	0.10	1270x
NOTCH1	P668S	0.32	0.32	1637x
NRAS	Q61K	0.12	0.14	1824x
PDGFRA	G426D	0.34	0.29	2018x
PIK3CA	E545K	0.09	0.16	773x
PIK3CA	H1047R	0.18	0.15	1694x

DNA from the HD200, a formalin-fixed cell line (Horizon Diagnostics) containing known variants, was evaluated using the TruSight Tumor 170 assay and sequenced on the NextSeq 500 System. 100% concordance was observed with expected frequency from all HD200 variants.

Table 4: Small Variant Detection with FFPE Tumor Samples

Sample	Reported Mutation	Detected Mutation	Detected Frequency	Coverage
FFPE_Colon	TP53 R158C	TP53 R158C	0.057	1545x
FFPE_Bone	TP53 P72R	TP53 P72R	0.059	515x
FFPE_Brain1	PIK3CA E545G	PIK3CA E545G	0.078	289x
FFPE_Brain2	PIK3CA H1047R	PIK3CA H1047R	0.076	531x
FFPE_Breast	KRAS G12D	KRAS G12D	0.049	1671x
FFPE_Lung1	KRAS G12D	KRAS G12D	0.059	575x
FFPE_Lung2	TP53 C242F	TP53 C242F	0.080	691x
FFPE_Skin	TP53 R248Q	TP53 R248Q	0.050	1240x

DNA from FFPE tumor samples was extracted and evaluated using the TruSight Tumor 170 assay and sequenced on the NextSeq 500 System. All 8 FFPE samples had 100% concordance with reported mutations.

Reliable Calling of Amplifications, Fusions, and Splice Variants from FFPE Samples

TruSight Tumor 170 combines the sensitivity of Illumina sequencing systems with new software platforms to enable simultaneous calling of amplifications, fusions, and splice variants. The TruSight Tumor 170 App features novel variant calling algorithms that produce accurate calls for splice variants, fusions, and gene amplifications from raw sequencing data in samples of varying quality (Tables 5 and 6).

Table 5: Amplification Calling with FFPE Tumor Samples

Sample	Reported Amplification	Reported Amplification Level	Detected Amplification	Detected Amplification Level
FFPE_Bone	FGF19	1.4	FGF19	2.9
FFPE_Brain2	PDGFRA	2.3	PDGFRA	2.9
FFPE_Breast	RPS6KB1	2.4	RPS6KB1	2.4
FFPE_Colon	BRCA2	2.2	BRCA2	2.0
FFPE_Lung1	PIK3CA	2.4	PIK3CA	2.7
FFPE_Lung2	FGFR1	2.4	FGFR1	2.9
FFPE_Lung3	MYC	2.2	MYC	2.8
FFPE_Lung4	CCNE1	2.1	CCNE1	2.2
FFPE_Lung5	EGFR	2.2	EGFR	4.5
FFPE_Lung6	CCND1	2.3	CCND1	2.9
FFPE_Stomach1	CDK6	2.3	CDK6	1.7
FFPE_Stomach2	MET	1.5	MET	1.4

DNA from FFPE tumor samples was extracted and then evaluated using the TruSight Tumor 170 assay and sequenced on the NextSeq 500 System. All 12 FFPE samples had 100% variant concordance.

Table 6: Fusion and Splice Variants Calling with FFPE Tissues and Cell Lines

Sample	DV200	Reported Variant	Detected Variant
FFPE_Brain Tissue	N/A	EGFR VIII Splice Variant	EGFR VIII Splice Variant
FFPE_Breast Tissue	81	RPS6KB1-VMP1, RPS6KB1-DIAPH3, CCDC170-ESR1 fusions	RPS6KB1-VMP1, RPS6KB1-DIAPH3, CCDC170-ESR1 fusions
FFPE_Ewing's Tissue	48.9	EWSR1-FLI1 fusion	EWSR1-FLI1 fusion
FFPE_Gastric Cell Line	93	MET Exon 14 Skipping Splice Variant	MET Exon 14 Skipping Splice Variant
FFPE_Lung CellLine	93	CCDC6-RET fusion	CCDC6-RET fusion
FFPE_Lung Tissue1	73.3	EML4-ALK fusion	EML4-ALK fusion
FFPE_Lung Tissue2	95	FGFR3-TACC3 fusions	FGFR3-TACC3 fusions
FFPE_Prostate Cell Line	95.5	ARv7 Splice Variant	ARv7 Splice Variant
FFPE_Prostate Tissue	28.7	TMPRSS2-ERG, TMPRSS2-GNPT fusions	TMPRSS2-ERG, TMPRSS2-GNPT fusions

RNA from FFPE tumor samples was extracted and then evaluated using the TruSight Tumor 170 assay and sequenced on the NextSeq 500 System. All 9 FFPE samples had 100% variant concordance. DV200 value is used to assess the quality of RNA used to prepare sequencing libraries, and represents the percentage of RNA fragments > 200 nucleotides.

Summary

TruSight Tumor 170 offers an integrated workflow solution for the detection of common somatic variants found in solid tumors. DNA and RNA libraries are prepared, sequenced, and analyzed simultaneously for efficient assessment of numerous types of somatic variants. Developed according to evidence-based guidelines, with input from key opinion leaders and late-stage pharmaceutical research, the panel provides labs with a comprehensive view of cancer-relevant genes and accurate analysis of low-frequency variants from FFPE DNA and RNA. By assessing 170 genes, and several different types of variants in a single assay, TruSight Tumor 170 offers a comprehensive genetic investigation of tumor samples in a streamlined solution.

Learn More

For more information about TruSight Tumor 170, visit www.illumina.com/TruSightTumor170

References

1. American Cancer Society. www.cancer.org. Accessed October 17, 2017.
2. O'Leary NA, Wright MW, Brister JR, et al. [Reference sequence \(RefSeq\) database at NCBI: current status, taxonomic expansion, and functional annotation](#). *Nucleic Acids Res.* 2016;44(D1):D733-45.
3. National Comprehensive Cancer Network. www.nccn.org. Accessed October 17, 2017.
4. European Society for Medical Oncology. www.esmo.org. Accessed October 17, 2017.

Ordering Information

Library Prep Kits	No. of Samples	Catalog No.
TruSight Tumor 170 Kit	24	OP-101-1004
TruSight Tumor 170 Kit, plus PierianDx	24	20032628
TruSight Tumor 170 Kit, with NextSeq v2.5 Reagents	24	20028821
TruSight Tumor 170 Kit, with NextSeq v2.5 Reagents, plus PierianDx	24	20032629

illumina, Inc. • 1.800.809.4566 toll-free (US) • +1.858.202.4566 tel • techsupport@illumina.com • www.illumina.com

© 2019 Illumina, Inc. All rights reserved. All trademarks are the property of Illumina, Inc. or their respective owners. For specific trademark information, see www.illumina.com/company/legal.html. Pub. No. 1170-2016-017-E QB5141

