

## Certificate of Analysis

### Description

---

<b>Product Name</b>	TruSeq Custom Amplicon Dx FFPE QC Box 1/1	<b>Catalog Number</b>	20006259
<b>Part Number</b>	20003464	<b>Lot Number</b>	A122818-1

### Test Conditions

The kitted reagents were tested for functional performance by qPCR amplification using control DNA (ACD1) and the included QC Primers (QCP).

### Test Results

Metric	Specification	Result
ACD1 Concentration (ng/ $\mu$ l)	$\geq 10$	Pass
Slope of ACD1 standard Curve	$-3.8 \leq X \leq -3.2$	Pass
R2 of ACD1 Standard Curve	$\geq 0.985$	Pass
NTC	$\geq 35.0$	Pass

### Certification

This document certifies that the product(s) described above meet quality specifications.

### Quality Review

Print Name : Kathy Phan

Signature



Date : 22-AUG-2019