

Business Stream Products
Certification Services

Date : 07/26/2017

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Attn: Karen Hui

Re. : CU US + Canada Certificate

Type of Equipment : Genetic Sequencer
Model Designation : See Certificate
Certificate No. : CU 72171865 0001
File No. : 31181842 032
Engineer/Contact : Jessamyn Duterte
Standard(s) : UL 61010-1:2012
UL 61010-2-081:2015
UL 61010-2-101:2015
CAN/CSA-C22.2 NO. 61010-1-12
CAN/CSA-C22.2 NO. 61010-2-081:15
CAN/CSA-C22.2 NO. 61010-2-101:15

Dear Ms. Hui,


The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72171865 0001. It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

Sincerely yours,
Certification Body


Dipl.-Ing. M. Glagla
QA Certification Officer

TUV Rheinland
of North America, Inc.
North American Headquarters

12 Commerce Road
Newtown, CT 06470

Tel 203-426-0888
Fax 203-426-4009
Toll Free TUV-RHEINLAND
Mail info@tuv.com
Web www.tuv.com

Member of
TUV Rheinland Group

Enclosure

Certificate



Certificate no.

CU 72171865 01

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

Illumina, Inc.
25861 Industrial Blvd.
Hayward CA 94545
USA

Test report no.: USA-JD 31181842 032

Client Reference: Karen Hui

Tested to:

UL 61010-1:2012
UL 61010-2-081:2015
UL 61010-2-101:2015
CAN/CSA-C22.2 NO. 61010-1-12
CAN/CSA-C22.2 NO. 61010-2-081:15
CAN/CSA-C22.2 NO. 61010-2-101:15

Certified Product: Genetic Sequencer

License Fee - Units

Model Designation: MiSeqDx

7

Rated Voltage: AC 100-240V, 50/60Hz
Rated Power: 400W
Protection Class: I

Appendix: 1, 1-6

7

Licensed Test mark:

Date of Issue

(day/mo/yr)

26/07/2017

