

Date : 03/03/2017

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Attn: Carol Escano

Re. : CU US + Canada Certificate

Type of Equipment : Next Generation Sequencer
Model Designation : See Certificate
Certificate No. : CU 72170569 0001
File No. : 31790423 004
Engineer/Contact : Ernesto Fernandez
Standard(s) : UL 61010-1:2012
UL 61010-2-010:2015
UL 61010-2-081:2015
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2
CAN/CSA-C22.2 NO. 61010-2-010:15
CAN/CSA-C22.2 NO. 61010-2-081:15

Dear Ms. Escano,

The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72170569 0001. It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

Sincerely yours,
Certification Body


Dipl.-Ing. M. Magla
QA Certification Officer

TUV Rheinland
of North America, Inc.
North American Headquarters

12 Commerce Road
Newtown, CT 06470

Tel 203-426-0888
Fax 203-426-4009
Toll Free TUV-RHEINLAND
Mail info@tuv.com
Web www.tuv.com

Member of
TUV Rheinland Group

Enclosure

Certificate



Certificate no.

CU 72170569 01

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

Illumina, Inc.
25861 Industrial Blvd.
Hayward CA 94545
USA

Test report no.: USA-EF 31790423 004

Client Reference: Carol Escano

Tested to:

UL 61010-1:2012
UL 61010-2-010:2015
UL 61010-2-081:2015
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2
CAN/CSA-C22.2 NO. 61010-2-010:15
CAN/CSA-C22.2 NO. 61010-2-081:15

Certified Product: Next Generation Sequencer

License Fee - Units

Model Designation: NovaSeq 6000

7

Rated Voltage: AC 200-240V, 50/60Hz
Rated Current: 15A (2500W)
Protection Class: I

Special Remarks: Replaces Certificate CU72170367.

7

Appendix: 1, 1-7

Licensed Test mark:

Date of Issue

(day/mo/yr)

02/03/2017



Certificate



Certificate no.

CU 72170569 02

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

VM Services Inc.
6723 Mowry Ave
Newark CA 94560
USA

Test report no.: USA-EF 31790423 004

Client Reference: Carol Escano

Tested to:

UL 61010-1:2012
UL 61010-2-010:2015
UL 61010-2-081:2015
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2
CAN/CSA-C22.2 NO. 61010-2-010:15
CAN/CSA-C22.2 NO. 61010-2-081:15

Certified Product: Next Generation Sequencer

License Fee - Units

Additional Manufacturing Plant: see above

(K756208)

Licensed Test mark:

Date of Issue

(day/mo/yr)

02/03/2017

