

## Certificate of Analysis

### Description

|              |   |                |          |
|--------------|---|----------------|----------|
| Product Name | TruSeq Custom Amplicon Dx FFPE QC Box 1/1 | Catalog Number | 20006259 |
| Part Number  | 20003464                                  | Lot Number     | A114733  |

### Test Conditions

The kitted reagents were tested for functional performance by qPCR amplification using control DNA (ACD1) and the included QC Primers (QCP).


### Test Results

| Metric                       | Specification           | Result |
|------------------------------|-------------------------|--------|
| ACD1 Concentration (ng/μl)   | ≥ 10                    | Pass   |
| Slope of ACD1 standard Curve | $-3.8 \leq X \leq -3.2$ | Pass   |
| R2 of ACD1 Standard Curve    | ≥ 0.985                 | Pass   |
| NTC                          | ≥ 35.0                  | Pass   |

### Certification

This document certifies that the product(s) described above meet quality specifications.

### Quality Review

|            |                |           |  |      |             |
|------------|----------------|-----------|--|------|-------------|
| Print Name | DONNA MAE CRUZ | Signature |  | Date | 23 MAR 2018 |
|------------|----------------|-----------|--|------|-------------|